




Certificate of Analysis

SALSA® FFPE Solution (SFS)

Catalogue #	SMR05	
Product name	SALSA® FFPE Solution (SFS)	
LOT	A1-0919	
	50 reactions.	
Shipping conditions	Dry ice or cooling elements.	
	Store upon arrival between -25°C and -15°C.	
	Expiration date: September 2022, when stored at recommended conditions.	
Purpose	The SALSA FFPE Solution (SFS) is used to extract DNA from formalin-fixed paraffin-embedded (FFPE) tissue samples.	
Quality control specifications	<ul style="list-style-type: none"> - Tests with previous and new lot with varying amounts of SFS: variation within the reference sample population and within the sample population is less than 10% (maximum stdv 0.10). All probes have a ratio between 0.8-1.2. - Conductivity measurements of previous and new lot: the difference is less than 10%. - Tests performed on Tris-EDTA (noDNA) samples: only the 64, 70, 76 and 82 nt Q-fragments are clearly visible. - Tests performed on 5 ng DNA are used to determine the variation of the primer peaks. This variation should be $\leq 20\%$. 	Test result
		PASS

This lot was certified by MRC Holland on 15 September 2020.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

More information: www.mrcholland.com ; www.mrcholland.eu	
	MRC Holland bv; Willem Schoutenstraat 1 1057 DL, Amsterdam, The Netherlands
E-mail	info@mrcholland.com (information & technical questions) order@mrcholland.com (orders)
Phone	+31 888 657 200

Implemented changes in the COA
Version 01 – 29 March 2021 (1) - Not applicable, new document.